



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/630,402	07/30/2003	Patrice Flaherty	1066	9003

7590  
R. Keith Harrison  
2139 E. Bert Kouns  
Shreveport, LA 71105

07/19/2006

EXAMINER
----------

HOEKSTRA, JEFFREY GERBEN

ART UNIT	PAPER NUMBER
----------	--------------

3736

DATE MAILED: 07/19/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

10/630,402

Applicant(s)

FLAHERTY, PATRICE

Examiner

Jeffrey G. Hoekstra

Art Unit

3736

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 26 June 2006.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-32 is/are pending in the application.
- 4a) Of the above claim(s) 12-23 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-11 and 24-32 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_

## DETAILED ACTION

### ***Notice of Amendment***

1. In response to the amendment filed on 06/26/2006, amended claims 1, 7, and 24 are acknowledged. The current rejections of the claims 1-11 and 24-32 are *withdrawn*. The following new and reiterated grounds of rejection are set forth:

### ***Claim Rejections - 35 USC § 102***

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

3. Claims 1-11 and 24-32 are rejected under 35 U.S.C. 102(b) as being anticipated by Strittmatter (US 5,396,899).

4. For claims 1, 7, and 24, Strittmatter discloses a multi-channel bodily-fluid handling device, comprising:

- a main tubing segment (11) for the passage of bodily fluids;
- an indicator unit (25) and a syringe port (24) disposed in fluid communication with said main tubing segment and in a branched relationship to each other, wherein said indicator unit has a fluid volumeter (30);
- a clamp (20) operably engaging said main tubing segment for selectively blocking fluid; and

Art Unit: 3736

- an air-permeable liquid-impervious membrane (34) disposed in said indicator unit.
5. For claims 2, 4, and 8, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a fluid volumeter (30) disposed in said indicator unit.
6. For claims 3, 6, 9, and 11, Strittmatter discloses a multi-channel bodily-fluid handling device wherein said indicator unit is disposed in removable fluid communication with said main tubing segment as best seen in Figures 5 and 6.
7. For claims 5, 10, and 26, Strittmatter discloses a multi-channel bodily-fluid handling device wherein said fluid volumeter is a volumeter chamber (column 4 lines 24-41).
8. For claim 25, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a connector (12) disposed in fluid communication with said main tubing segment and disposed in removable fluid communication with said indicator unit.
9. For claim 27, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a collector conduit (21) disposed in fluid communication with said main tubing segment and disposed in fluid communication with said indicator unit.
10. For claim 28, Strittmatter discloses a multi-channel bodily-fluid handling device wherein said indicator unit comprises a volumeter conduit (19) disposed in fluid communication with said collection conduit and disposed in fluid communication with said volumeter conduit.
11. For claim 29, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a port (22) disposed between said collector and volumeter conduits.

Art Unit: 3736

12. For claim 30, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a syringe tubing (15) segment disposed in fluid communication with said main tubing segment and wherein said port (24) is disposed on said syringe tubing segment.

13. For claim 31, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a connector (13) disposed in said main tubing segment between said clamp and said indicator unit and said port.

14. For claim 32, Strittmatter discloses a multi-channel bodily-fluid handling device, further comprising: a port (32) in fluid communication with said indicator unit, an attachment sleeve (31) provided in fluid communication with said port, and a fluid receptacle (31) in fluid communication with said attachment sleeve.

### ***Response to Arguments***

15. Applicant's arguments with respect to claims 1-11 and 24-32 have been considered but are moot in view of the new ground(s) of rejection.

### ***Conclusion***

16. The following prior art made of record and not relied upon is considered pertinent to applicant's disclosure: Dieffenbach (US 5,089,421) discloses a bodily-fluid collection apparatus with air-permeable liquid-impervious membrane.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jeffrey G. Hoekstra whose telephone number is


Art Unit: 3736

(571)272-7232. The examiner can normally be reached on Monday through Friday, 8:00 a.m. to 5:00 p.m. EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max F. Hindenburg can be reached on (571)272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

JGH



MAX F. HINDENBURG  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 3700